



EMBASSY OF THE
UNITED STATES OF AMERICA
P.O. BOX 606 - 00621, VILLAGE MARKET
NAIROBI, KENYA

August 7th, 2019

RE: INVITATION TO BID ON PR8486568: MRU-SUPPLY, INSTALLATION & COMMISSIONING OF AUTOCLAVES IN MRD SITES (4 EACH)

The American Embassy Nairobi, Kenya, has a requirement for Supply, Installation & Commissioning of Autoclaves in MRU sites mentioned below:-

1. **CRC Labs Kericho (2 each)**
2. **Respiratory Lab Nairobi (1 each)**
3. **Microbiology lab Nairobi (1 each)**

Your proposal must be submitted in two separate attachments addressed to: To the Contracting Officer, American Embassy, through GitariRW@state.gov and NairobiGSO-bids@state.gov on or before **August 19th 2019 – 1000hrs** and should include:-

- **Cost of Supply, Installation and Commissioning of Autoclave with accessories with 1 year warranty. CAT#5075ELV-WR, External Dimensions (WxHxD (mm))-739x1075x800; Chamber Volume (Liter)-160**
- **After sale service/warranty for 1 year from the date of Installation**
- **Must Attach Certificate of Origin when delivering the items**
- **Power Ratings: 240v, 50 Hz**
- **List all the accessories, elements, specifications that come with the item.**
- **Mobile lifting device for loading and unloading the autoclaves**
- **Biohazard exhaust filtration system**
- **Condensate collection vessel, 30 liter**

The subject heading should be as follows:

1. *RFQ NO. ON PR8486568 PRICE PROPOSAL (Must have the Company Letter Head or Company Stamp/Seal including your quotation Reference Number).*
2. *RFQ NO. ON PR8486568 TECHNICAL QUALIFICATION Proposal (Must not have your Letter Head or Company Stamp/Seal i.e. should be on plain paper). Your quotation Reference Number is required on this which matches the price proposal number.*

The U.S. Government intends to award a contract/purchase order to a responsible company submitting an acceptable quotation at the reasonable price. The RFQ does not commit the American Embassy to make any award. The Embassy may cancel this RFQ or any part of it.

Sincerely,

A handwritten signature in black ink, appearing to read 'NT', with a long horizontal stroke extending to the right.

Nathaniel Tishman
Contracting Officer